ABL inc.
ADVANCED BIOSCIENCE LABORATORIES

CERTIFICATE OF ANALYSIS

Product:	HIV-1 p24 Antigen Capture Assay
Cat#:	5421 and 5447
Lot#:	2174
Release Date:	10/14/2024
Expiry Date:	SEP2025

HIV-1 p24 Antigen Capture Assay kit of this lot consists of the following:

Disruption Buffer	Lot# DSB141
Microelisa Plate, Anti-HIV-1 p24	Lot# 24081502
HIV-1 p24 Standard	Lot# P24STD113
Conjugate Solution	Lot# P24C164
Peroxidase Substrate	Lot# 231012
Stop Solution (2N H_2SO_4)	Lot# 2407J57
Wash Buffer (20X)	Lot# 021224-04
Plate Sealers	Lot# 20240398

Storage condition for HIV-1 p24 Antigen Capture Assay is 2-8°C.

	Specifications	**Results**
Absorbance of 100 pg/ml Standard Antigen	1.2 to 2.2	1.923
Absorbance of Negative Control	< 0.120	0.068
Linear Regression R squared	> 0.995	0.9989

This lot of P24 Elisa kit has been tested and meets Quality Control testing requirements of archived curve and Negative control results

Peter Silvera 14 Oct 2024

Peter Silvera Date
Program Director III, Innovation & Translation

Advanced BioScience Laboratories, Inc.
9800 Medical Center Drive – Building D
Rockville, MD 20850
Phone : 1-301-816-5459
Fax : 1-301-816-5438
info@ablinc.com
www.ablinc.com