ABL inc.
ADVANCED BIOSCIENCE LABORATORIES

CERTIFICATE OF ANALYSIS

Product:	HIV-1 p24 Antigen Capture Assay
Cat#:	5421 and 5447
Lot#:	2183
Release Date:	06/25/2026
Expiry Date:	MAY 2027

HIV-1 p24 Antigen Capture Assay kit of this lot consists of the following:

Disruption Buffer	Lot# DSB149
Microelisa Plate, Anti-HIV-1 p24	Lot# 26040901
HIV-1 p24 Standard	Lot# P24STD120
Conjugate Solution	Lot# P24C174
Peroxidase Substrate	Lot# 250806
Stop Solution (2N H_2SO_4)	Lot# 2510J00
Wash Buffer (20X)	Lot# 041326-04
Plate Sealers	Lot# 20250592

Storage condition for HIV-1 p24 Antigen Capture Assay is 2-8°C.

	Specifications	**Results**
Absorbance of 100 pg/ml Standard Antigen	1.2 to 2.2	1.6
Absorbance of Negative Control	< 0.120	0.08
Linear Regression R squared	> 0.995	0.999

This lot of P24 Elisa kit has been tested and meets Quality Control testing requirements of archived curve and Negative control results

Brian Taylor 01JUL2026

Brian Taylor Date
Program Director III, Innovation & Translation

Advanced BioScience Laboratories, Inc.
9800 Medical Center Drive – Building D
Rockville, MD 20850
Phone : 1-301-816-5459
Fax : 1-301-816-5438
info@ablinc.com
www.ablinc.com