ABL inc.

ADVANCED BIOSCIENCE LABORATORIES

CERTIFICATE OF ANALYSIS

Product:	SIV p27 Antigen Capture Assay
Cat#:	5436 and 5450
Lot#:	S2792
Release Date:	12/12/19
Expiry Date:	11/2020

SIV p27 Antigen Capture Assay kit of this lot consists of the following:

Microelisa Plate, Anti-SIV-p27	Lot# 19020701
SIV p27 Standard	Lot# P27STD27
Disruption Buffer	Lot# DB1085
Conjugate Solution	Lot# P27C69
Peroxidase Substrate	Lot# 190313
Stop Solution (2N H_2SO_4)	Lot# 4906907
Wash Buffer (20X)	Lot# 020719-01
Plate Sealers	Lot# 16119654

Storage condition for SIV p27 Antigen Capture Assay is 2-8°C.

	Specifications	**Results**
Absorbance of 2000 pg/ml Standard Antigen	1.5 to 2.0	1.5 OD
Determined p27 Control Antigen Concentration	900 to 1100 pg/ml	928 pg/ml
Absorbance of Negative Control	< 0.100	0.073 OD
Linear Regression R squared	> 0.990	1.000

This product conforms to the specifications listed above.

12/12/19

Stephen C. Whitney — Date
Laboratory Manager, Immunology

Advanced BioScience Laboratories, Inc.
9800 Medical Center Drive – Building D
Rockville, MD 20850
Phone : 1-301-816-5459
Fax : 1-301-816-5438
info@ablinc.com
www.ablinc.com