

CERTIFICATE OF ANALYSIS

Product: HIV-1 p24 Antigen Capture Assay
Cat#: 5421 and 5447
Lot#: 2161B
Release Date: 04/07/21
Expiry Date: 03/2022


HIV-1 p24 Antigen Capture Assay kit of this lot consists of the following:

Disruption Buffer	Lot# DSB128
Microelisa Plate, Anti-HIV-1 p24	Lot# 20130501
HIV-1 p24 Standard	Lot# P24STD91B
Conjugate Solution	Lot# P24C151
Peroxidase Substrate	Lot# 201116
Stop Solution (2N H ₂ SO ₄)	Lot# 4009E68
Wash Buffer (20X)	Lot# 121020-01
Plate Sealers	Lot# 14320294

Storage condition for HIV-1 p24 Antigen Capture Assay is 2-8°C.

	Specifications	Results
Absorbance of 100 pg/ml Standard Antigen	1.4 to 2.0	1.5
Determined p24 Control Antigen Concentration	45 to 55 pg/ml	52 pg/ml
Absorbance of Negative Control	< 0.10	0.09
Linear Regression R squared	> 0.995	0.998

This product conforms to the specifications listed above.



Stephen Whitney
Laboratory Manager

04/07/21

Date