



ADVANCED BIOSCIENCE LABORATORIES

CERTIFICATE OF ANALYSIS

Product: HIV-1 p24 Antigen Capture Assay
Cat#: 5421 and 5447
Lot#: 2162
Release Date: 06/23/21
Expiry Date: 05/2022

HIV-1 p24 Antigen Capture Assay kit of this lot consists of the following:

Disruption Buffer Lot# DSB129
Microelisa Plate, Anti-HIV-1 p24 Lot# 21042301
HIV-1 p24 Standard Lot# P24STD94
Conjugate Solution Lot# P24C152
Peroxidase Substrate Lot# 210317
Stop Solution (2N H2SO4) Lot# 4104604
Wash Buffer (20X) Lot# 040621-01
Plate Sealers Lot# 29520488

Storage condition for HIV-1 p24 Antigen Capture Assay is 2-8°C.

Table with 3 columns: Specifications, Results. Rows include Absorbance of 100 pg/ml Standard Antigen (1.4 to 2.0, 1.8), Determined p24 Control Antigen Concentration (45 to 55 pg/ml, 49 pg/ml), Absorbance of Negative Control (< 0.10, 0.08), Linear Regression R squared (> 0.995, 1.000)

This product conforms to the specifications listed above.

Signature of Stephen Whitney and date 06/23/21