



ADVANCED BIOSCIENCE LABORATORIES

CERTIFICATE OF ANALYSIS

Product: HIV-1 p24 Antigen Capture Assay
Cat#: 5421 and 5447
Lot#: 2157
Release Date: 07/16/20
Expiry Date: 03/2021


HIV-1 p24 Antigen Capture Assay kit of this lot consists of the following:

Disruption Buffer	Lot# DSB125
Microelisa Plate, Anti-HIV-1 p24	Lot# 20042401
HIV-1 p24 Standard	Lot# P24STD87
Conjugate Solution	Lot# P24C147B
Peroxidase Substrate	Lot# 181114
Stop Solution (2N H ₂ SO ₄)	Lot# 4906907
Wash Buffer (20X)	Lot# 112719-02
Plate Sealers	Lot# 33919950

Storage condition for HIV-1 p24 Antigen Capture Assay is 2-8°C.

	Specifications	Results
Absorbance of 100 pg/ml Standard Antigen	1.4 to 2.0	1.6 OD
Determined p24 Control Antigen Concentration	45 to 55 pg/ml	47 pg/ml
Absorbance of Negative Control	< 0.10	0.07
Linear Regression R squared	> 0.995	0.999

This product conforms to the specifications listed above.

 07/16/20

Stephen Whitney Date
Laboratory Manager