



ADVANCED BIOSCIENCE LABORATORIES

CERTIFICATE OF ANALYSIS

Product: HIV-1 p24 Antigen Capture Assay
 Cat#: 5421 and 5447
 Lot#: 2165
 Release Date: 05/11/2022
 Expiry Date: Apr2023


HIV-1 p24 Antigen Capture Assay kit of this lot consists of the following:

Disruption Buffer	Lot# DSB132
Microelisa Plate, Anti-HIV-1 p24	Lot# 22040101
HIV-1 p24 Standard	Lot# P24STD104
Conjugate Solution	Lot# P24C155
Peroxidase Substrate	Lot# 210317
Stop Solution (2N H ₂ SO ₄)	Lot# 4104G04
Wash Buffer (20X)	Lot# 042722-01
Plate Sealers	Lot# 19321155

Storage condition for HIV-1 p24 Antigen Capture Assay is 2-8°C.

	Specifications	Results
Absorbance of 100 pg/ml Standard Antigen	1.4 to 2.0	1.7
Determined p24 Control Antigen Concentration	45 to 55 pg/ml	46 pg/ml
Absorbance of Negative Control	< 0.100	0.070
Linear Regression R squared	> 0.995	1.000

This product conforms to the specifications listed above.



 Irene Kalisz
 Staff Scientist

5/16/22

 Date