



ADVANCED BIOSCIENCE LABORATORIES

CERTIFICATE OF ANALYSIS

Product: HIV-1 p24 Antigen Capture Assay
 Cat#: 5421 and 5447
 Lot#: 2169
 Release Date: 06/26/2023
 Expiry Date: MAY2024

HIV-1 p24 Antigen Capture Assay kit of this lot consists of the following:

Disruption Buffer	Lot# DSB136
Microelisa Plate, Anti-HIV-1 p24	Lot# 23060901
HIV-1 p24 Standard	Lot# P24STD108
Conjugate Solution	Lot# P24C159
Peroxidase Substrate	Lot# 220929
Stop Solution (2N H ₂ SO ₄)	Lot# 2106B48
Wash Buffer (20X)	Lot# 060623-02
Plate Sealers	Lot# 28422136

Storage condition for HIV-1 p24 Antigen Capture Assay is 2-8°C.

	Specifications	Results
Absorbance of 100 pg/ml Standard Antigen	1.4 to 2.0	1.5
Determined p24 Control Antigen Concentration	45 to 55 pg/ml	45 pg/ml
Absorbance of Negative Control	< 0.100	0.060
Linear Regression R squared	> 0.995	1.000

This product conforms to the specifications listed above.


 Stephen Jackson
 Head, Immunology Operations

13-Jul-2023
 Date