



ADVANCED BIOSCIENCE LABORATORIES

CERTIFICATE OF ANALYSIS

Product: SIV p27 Antigen Capture Assay
 Cat#: 5436 and 5450
 Lot#: S2793C
 Release Date: 06/11/20
 Expiry Date: 05/2021


SIV p27 Antigen Capture Assay kit of this lot consists of the following:

Microelisa Plate, Anti-SIV-p27	Lot# 19083001
SIV p27 Standard	Lot# P27STD27
Disruption Buffer	Lot# DB1086
Conjugate Solution	Lot# P27C71C
Peroxidase Substrate	Lot# 181114
Stop Solution (2N H ₂ SO ₄)	Lot# 4906907
Wash Buffer (20X)	Lot# 082019-02
Plate Sealers	Lot# 33919950

Storage condition for SIV p27 Antigen Capture Assay is 2-8°C.

	Specifications	Results
Absorbance of 2000 pg/ml Standard Antigen	1.5 to 2.0	1.7 OD
Determined p27 Control Antigen Concentration	900 to 1100 pg/ml	1009 pg/ml
Absorbance of Negative Control	< 0.100	0.059 OD
Linear Regression R squared	> 0.990	1.000

This product conforms to the specifications listed above.

 10/19/20

 Stephen C. Whitney Date
 Laboratory Manager, Immunology